

Instrument

Centrifuge (x3)

Model / Number

5424

Manufacturer

Eppendorf

Location

Life Science 250

**Instrument Description**

24 x 1.5/2mL; Max speed/RCF 15,000/21,130xg, 100rpm increments, aerosol tight lid, 4 x 8 PCR tube rotor available in lab for this model.

User Charges

Not applicable

User Information

User must have key card access to Life Science 250. User must be trained to use this equipment. The user is responsible for balancing the load, cleaning and performing per run maintenance, completing the centrifuge log and providing any disposable supplies required for operation. Any damage caused by the user is the responsibility of the user to repair or replace.